



**INTEK, INC.**

**RHEOTHERM<sup>®</sup> AND RHEOVAC<sup>®</sup> INSTRUMENTS**

751 Intek Way • Westerville, OH 43082-9057  
Tel: (614)895-0301 • Fax: (614)895-0319  
www.intekflow.com • sales@intekflow.com

Certified to  
ISO 9001:2000

## DECLARATION OF CONFORMITY

**For the Rheotherm Flow Instruments:**

**Model 210-I & Model 210-XDT  
(Excludes Model 210R option)**

is herewith confirmed to comply with the requirements set out in the Council Directive on the Approximation of Laws of the Member States relating to Electromagnetic Compatibility Directive 89/336/EEC. The following standards were applied:

- |              |  |
|--------------|--|
| EN61000-6-4; | Electromagnetic compatibility generic <i>emission</i> standard Part 2: Industrial environment.   |
| EN61000-6-2; | Electromagnetic compatibility generic <i>immunity</i> standard Part 2: Industrial environment.   |
| EN 55011;    | Limits and methods of measuring of radio disturbance characteristics of Industrial, Scientific, and Medical (ISM) radio-frequency equipment - Class A, group 1 equipment limits of EN 55011. |

Manufactured by:

Intek, Inc.  
751 Intek Way  
Westerville, OH 43082  
United States of America

Signed this 18 day of July, 2005

Joseph W. Harpster  
President, Intek, Inc.